



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R60/2000-GB1-04.05

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue, Teddington**
Middlesex, TW11 0JZ
United Kingdom

Person responsible: **P Dixon – Business Team Manager Type Approval & Test**

Applicant Name: **Gicam S.N.C Di Carrara Danilo & Co**
Address: **Piazza XI Febbraio, 2**
22015 Gravedona (CO)
Italy

Manufacturer of the certified pattern is:

The applicant

Identification of the certified pattern:

Single-ended shear beam (bending) strain gauge load cell

Model Designation	TA-4				
Maximum capacity, E_{\max} (kg)	50	100	150	250	360
Minimum verification interval, V_{\min} (kg)	0.004	0.008	0.012	0.016	0.03
Accuracy class	C3				
Maximum number of load cell intervals, n_{\max}	3000				
Apportionment factor; p_{LC}	0.7				

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology -OIML):

R 60 Metrological regulation for load cells Edition: 2000 (E) for accuracy class: C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.